

Sensor Technical Specifications

Reyocto Agrimation Private Limited

3rd Floor, No.33, Sana Enclave, 17th Cross, MC Layout, Vijayanagar- Bangalore-560040.

Phone: +91 80 43711222. Email: contact@reyocto.com

Turbidity Sensor



Measuring principle	Scattering light method
Measuring range	0.01-100 NTU, 0.01 — 4000 NTU
Accuracy	Less than the measured value of $\pm 1\%$, or ± 0.1 NTU
Resolution	0.01NTU
Calibration method	Sample calibration, slope calibration
Working conditions	0~45℃, <0.4MPa
Output Method	RS-485(Modbus/RTU)
Source	12 VDC
Protection Level	IP68 / NEMA6P

Reyocto Agrimation Private Limited

3rd Floor, No.33, Sana Enclave, 17th Cross, MC Layout, Vijayanagar- Bangalore-560040.

Phone: +91 80 43711222. Email: contact@reyocto.com