

Sensor Technical Specifications

Reyocto Agrimation Private Limited

#11, Churchaghatta Kothanur Main Road, J.P. Nagar 7th Phase, Bangalore – 560 062

Phone: +91 80 43711222. Email: contact@reyocto.com

COD Sensor



Measuring range	0.75~500mg/L
Accuracy	<5%
Resolution	0.01mg/L
Calibration method	One or Two-point calibration
Temperature Range	+5~45°C
Output Method	RS-485(Modbus/RTU)
Protection Level	IP68
Installation Method	Immersion mounting, 3/4 NPT thread

Reyocto Agrimation Private Limited

#11, Churchaghatta Kothanur Main Road, J.P. Nagar 7th Phase, Bangalore – 560 062

Phone: +91 80 43711222. Email: contact@reyocto.com